सीएसआईआर–उत्तर पूर्व विज्ञान तथा प्रोद्योगिकी संस्थान, जोरहाट , असम CSIR - NORTH-EAST INSTITUTE OF SCIENCE AND TECHNOLOGY, JORHAT (ASSAM), INDIA

This is to inform that in the tender for "Supply and Installation of CHNS Analyzer" having Tender ID "2019_CSIR_39227_1", some points have been revised and the same may be read as under:

Revised Specifications for CHNS Analyzer

Earlier Specification	Revised may be read as -
Analysis Time:	
For all the elements the analysis should	For all the elements the analysis should be
be around ≤ 5 min simultaneously	around ≤ 15 min simultaneously
Electronic Balance:	
An electronic balance of high precision	An electronic balance of high precision
0.1 mg (four digits after decimal places)	0.001 – 0.01 mg (four digits after decimal
with automatic weight transfer facility	places) with automatic weight transfer
should be provided	facility should be provided

All other specifications, terms and conditions will remain same.