

**सीएसआईआर-उत्तर पूर्व विज्ञान तथा प्रौद्योगिकी संस्थान, जोरहाट , असम**  
**CSIR - NORTH-EAST INSTITUTE OF SCIENCE AND TECHNOLOGY,**  
**JORHAT (ASSAM), INDIA**

This is to inform that in the tender for “Supply and Installation of Bench Top High Sensitive Triple/Tandem Quadrupole LC-MS/MS System Connected with an UHPLC System” having Tender ID “2019\_CSIR\_39207\_1”, some points have been revised and the same may be read as under:

**Revised Specifications for Supply and Installation of Bench Top High Sensitive Triple/Tandem Quadrupole LC-MS/MS System Connected with an UHPLC System**

<b>Earlier Specification</b>	<b>Revised may be read as -</b>
<b><u>Ionization source</u></b> “Switching between ESI and APCI should be ≤ 25 ms”	<b>This point is deleted.</b>
<b><u>Operating modes</u></b> “Information dependent acquisition system or equivalent scan mode of MRM to high sensitivity product ion scan for library confirmation”	<b>This point is deleted.</b>
<b><u>UHPLC</u></b> “The carryover of the autosampler must be less than 0.002% or better”	This point is amended as “ <b>The carryover of the autosampler must be less than 0.005%.</b> ”
<b><u>UHPLC</u></b> “UHPLC consumables like injection syringes, tubings, mass standards, filters, etc. should be quoted for the entire warranty period.”	<b>This point is deleted</b>
<b><u>Other</u></b>	The following two points have been added additionally: <b>1. The genuine part numbers of major accessories should be available in supplier’s public website</b> <b>2. Petty consumables like injection syringes, tubings, gaskets, seals, spares, etc. should be quoted for the entire warranty period.</b> <b>3. The MS system should be compatible with other UHPLC systems.</b>
<b>All other specifications, terms and conditions will remain same.</b>	

This is also to inform that the following dates of the tender for “Supply and Installation of Bench Top High Sensitive Triple/Tandem Quadrupole LC-MS/MS System Connected with an UHPLC System” having Tender ID “2019\_CSIR\_39207\_1” have been extended :

**New Dates:**

Upload of Amendment, if any: extended to 22.01.2020

Last Date & time of Submission of bids: 05/02/2019 at 11:00 AM

Last Date & time of Opening of bids: 06/02/2019 at 11:00 AM