

सीएसआईआर—उत्तरपूर्वविज्ञानएवंप्रौद्योगिकीसंस्थान, जोरहाट - 785006 असम, भारत

**CSIR - NORTH-EAST INSTITUTE OF SCIENCE AND TECHNOLOGY,
JORHAT (ASSAM), INDIA**

This is to inform that in the technical specifications, the following amendments and modifications have been made in the tender for “**Supply and Installation of Bomb Calorimeter and Accessories**” having Tender ID “**2019_CSIR_33348_1**”. Prospective bidders may read the following points and quote accordingly.

Sl. No. as per NIT Specification	Specifications	Modified/may be read as
1	Operation	Earlier: Automatic operation; System should have facility for auto open and close provision for main cover. The earlier details is replaced as below:- “Automatic pre and post combustion system ; no external bucket; System main cover should open and close by automatic or manual. ”
3	Test Method	Earlier: Should comply with the following: ASTM D 240, ASTM D4809, ASTM D 5865, ISO 9831, ISO 1928, BIS 1350-part II and DIN GB The earlier details is replaced as below:- “ASTM D 240, ASTM D4809, ASTM D 5865, ASTM-E144, ISO 9831, ISO 1928, BIS 1350- part II and DIN GB”
14	Gas Filling Arrangement	Earlier: The system should have facility for Automatic gas filling with auto shut off arrangement after loading the bomb into the calorimeter. The earlier details is replaced as below:- “It should have Automatic gas filling with auto shut off arrangement.”

The tender submission end time and tender opening time is also extended further by one week as below :

Bid Submission End Time: 07.11.2019 11:00 AM

Bid Opening Date: 08.11.2019 11:00 AM

All other tendered specifications will remain un-changed.